

**T.C.L. Tradinco Calibration Laboratory**

**CALIBRATION CERTIFICATE**

Page 1 of 2

Certificate number : 22200792

Applicant  
Revin Brielle B.V.  
Industrieterrein 't Woud nr. 16  
3232 LN Brielle

Instrument  
Pressure transmitter  
Manufacturer  
Wika  
Type  
P-30  
Serial number  
2276854  
Service number  
18038  
Customer identification number  
-

Calibration method  
This calibration is carried out with :  
Deadweight tester      type : M2000/1      S/N : 7518/93      Cal. Due : 21 May 2024  
Digital Multimeter      type : 1281      S/N : 43394      Cal. Due : 10 October 2022

During calibration a test pressure is put under a calibrated measuring piston which is loaded with calibrated weights

The instrument is calibrated in combination with oil/gas separator type 8335.

Environmental conditions  
Temperature:      20.0°C ± 2°C  
Humidity:      55 %rh ± 25 %rh

Date / period of calibration  
17 June 2022

Results  
See following page(s)

Uncertainty  
± 0.020 % of full scale with use of correction figures in the range of 0 - 100 % of full scale


The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of  $k = 2$ , which provides a confidence level of approximately 95 %.  
The standard uncertainty has been determined in accordance with document EA-4/02.

Calculated by software revision: 2.7.6.6/2.45

Traceability  
The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the RvA.

Date  
17 June 2022

Name  
S. Baidjnath-Misier  
Head of TCL





T.C.L. Tradinco Calibration Laboratory

## CALIBRATION CERTIFICATE

Page 2 of 2

Certificate number : 22200792

	Standard	Standard	Instrument
Service number			18038
Type	M2000/1	1281	P-30
Manufacturer	Tradinco Instruments	Wavetek	Wika
Description	Deadweight tester	Digital Multimeter	Pressure transmitter
Serial number	7518/93	43394	2276854
Range	(2.6 ÷ 350) bar	(0 ÷ 100) mA	(0 ÷ 40) bar
Medium	Oil		Air
Accuracy	±0.011 % R	±0.018 % R	±0.1 % S
Temperature	20.58°C ± 0.5°C	20.58°C ± 0.5°C	

MP.	STANDARD		INSTRUMENT					
	INPUT UP	INPUT DOWN	READING UP	READING DOWN	ERROR UP	ERROR DOWN	LIMIT UP	LIMIT DOWN
	bar	bar	mA	mA	% S	% S	% S	% S
1	0.00000	0.00000	3.993	3.991	-0.04	-0.06	± 0.10	± 0.10
2	9.99605	9.99605	7.990	7.989	-0.05	-0.06	± 0.10	± 0.10
3	19.9965	19.9965	11.989	11.989	-0.06	-0.06	± 0.10	± 0.10
4	29.9968	29.9968	15.987	15.986	-0.07	-0.08	± 0.10	± 0.10
5	39.9907		19.985		-0.07		± 0.10	

The pressure transmitter is calibrated in vertical position.

The pressure transmitter has not been adjusted (As found is as left calibration).

Customer ID. no. :-

Customer ref. : 20220449

Our ref. : 62200051

Customer : Revin Brielle B.V.

Calibrated by  
Calibration Technician

R. Gafoer

Cal. date : 17 June 2022

Approved by  
Head of TCL

S. Baidinath-Misier

Date: 17 June 2022

Zoetermeer