

T.C.L. Tradinco Calibration Laboratory

CALIBRATION CERTIFICATE

Page 1 of 2

Certificate number : 22200791

Applicant Revin Brielle B.V.
 Industrieterrein 't Woud nr. 16
 3232 LN Brielle

Instrument Pressure transmitter
 Manufacturer Wika
 Type P-30
 Serial number 2276853
 Service number 22524
 Customer identification number -

Calibration method This calibration is carried out with :
 Deadweight tester type : M2000/1 S/N : 7518/93 Cal. Due : 21 May 2024
 Digital Multimeter type : 1281 S/N : 43394 Cal. Due : 10 October 2022

During calibration a testpressure is put under a calibrated measuring piston which is loaded with calibrated weights

The instrument is calibrated in combination with oil/gas separator type 8335.

Environmental conditions Temperature: 20.0°C ± 2°C
 Humidity: 55 %rh ± 25 %rh

Date / period of calibration 17 June 2022

Results See following page(s)


Uncertainty ± 0.020 % of full scale with use of correction figures in the range of 0 - 100 % of full scale

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor of $k = 2$, which provides a confidence level of approximately 95 %.
 The standard uncertainty has been determined in accordance with document EA-4/02.

Calculated by software revision: 2.7.6.6/2.45

Traceability The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the RvA.

Date 17 June 2022



Name S. Baidjnath-Misier
 Head of TCL

T.C.L. Tradinco Calibration Laboratory

CALIBRATION CERTIFICATE

Page 2 of 2

Certificate number : 22200791

	Standard	Standard	Instrument
Service number			22524
Type	M2000/1	1281	P-30
Manufacturer	Tradinco Instruments	Wavetek	Wika
Description	Deadweight tester	Digital Multimeter	Pressure transmitter
Serial number	7518/93	43394	2276853
Range	(2.6 ÷ 350) bar	(0 ÷ 100) mA	(0 ÷ 160) bar
Medium	Oil		Air
Accuracy	±0.011 % R	±0.018 % R	±0.1 % S
Temperature	20.3°C ± 0.5°C	20.3°C ± 0.5°C	

MP.	STANDARD		INSTRUMENT					
	INPUT UP bar	INPUT DOWN bar	READING UP mA	READING DOWN mA	ERROR UP % S	ERROR DOWN % S	LIMIT UP % S	LIMIT DOWN % S
1	0.00000	0.00000	4.003	4.003	0.02	0.02	± 0.10	± 0.10
2	39.9909	39.9909	8.003	8.004	0.02	0.03	± 0.10	± 0.10
3	79.9844	79.9844	12.002	12.002	0.02	0.02	± 0.10	± 0.10
4	119.977	119.977	16.001	16.001	0.02	0.02	± 0.10	± 0.10
5	159.968		19.998		0.01		± 0.10	

The pressure transmitter is calibrated in vertical position.
 The pressure transmitter has not been adjusted (As found is as left calibration).

Customer ID. no. : -
 Customer ref. : 20220449
 Our ref. : 62200051
 Customer : Revin Brielle B.V.

Calibrated by
 Calibration Technician
 R. Gafoer
 Cal. date : 17 June 2022

Approved by
 Head of TCL

 S. Baidinath-Misier
 Date: 17 June 2022
 Zoetermeer